

-----  
Midmark M11 - v1.0.0  
Total Cycles: 888

Name: RITTER 1  
Sterilizer ID: V2178198  
06/14/2021 07:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.5 F 0.0  
2:00 173.5 F 2.0  
4:00 186.2 F 6.0  
6:00 211.9 F 12.4  
8:00 226.3 F 7.1  
10:00 241.4 F 13.1  
12:00 241.0 F 10.9  
14:00 238.2 F 9.5  
16:00 253.7 F 17.0  
18:00 266.4 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 28.1  
1:00 272.6 F 28.6  
2:00 272.8 F 28.9  
3:00 271.8 F 28.2  
4:00 272.3 F 28.3  
  
Min 271.6 F 27.9  
Max 272.9 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.7 F 0.0  
Dry Cycle Aborted

FILLING: 1:23  
HEATING: 19:17  
STERILIZING: 4:00  
VENTING: 2:52  
DRYING: 0:21  
TOTAL CYCLE: 00:27:53

06/14/2021 08:09 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01