
Midmark M11 - v1.0.0
Total Cycles: 1627

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2022 10:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.0 F 0.0
2:00 180.3 F 2.4
4:00 178.2 F 5.9
6:00 209.4 F 11.3
8:00 220.0 F 5.6
10:00 232.2 F 10.7
12:00 234.7 F 7.2
14:00 244.6 F 13.4
16:00 243.6 F 11.9
18:00 256.2 F 18.5
20:00 267.4 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.2
1:00 272.3 F 28.8
2:00 271.3 F 28.3
3:00 272.4 F 28.8
4:00 271.3 F 28.1

Min 271.2 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.1 F 0.0
5:00 232.0 F 0.0
10:00 227.1 F 0.0
15:00 226.9 F 0.0
20:00 227.5 F 0.0
25:00 229.1 F 0.0
30:00 230.8 F 0.0
Drying Complete

FILLING: 1:09
HEATING: 21:25
STERILIZING: 4:00
VENTING: 3:03
DRYING: 30:00
TOTAL CYCLE: 00:59:36
CYCLE COMPLETE

02/14/2022 11:21 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____