
Midmark M11 - v1.0.5
Total Cycles: 1047

Name: RITTER2
Sterilizer ID: V2313228
04/28/2022 11:47 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 127.9 F 0.0
2:00 128.8 F 1.6
4:00 146.5 F 4.3
6:00 171.2 F 8.3
8:00 194.0 F 13.0
10:00 213.2 F 6.2
12:00 226.5 F 10.9
14:00 229.7 F 7.3
16:00 240.9 F 12.9
18:00 240.8 F 11.4
20:00 252.8 F 17.6
22:00 263.8 F 24.1

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 29.4
1:00 272.2 F 30.0
2:00 271.6 F 29.2
3:00 272.3 F 29.9
4:00 271.5 F 29.3

Min 271.4 F 29.2
Max 272.6 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.1 F 0.0
5:00 209.2 F 0.0
10:00 222.0 F 0.0
15:00 228.4 F 0.0
20:00 224.0 F 0.0
25:00 222.1 F 0.0
30:00 220.2 F 0.0
Drying Complete

FILLING: 1:02
HEATING: 23:58
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:02:11
CYCLE COMPLETE

04/28/2022 12:49 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____