
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 003002
14:01 20/01/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:56
PRESSURIZING 2:58
STERILIZING 3:54
136.5 C 224kPa 3:54
136.4 C 220kPa 4:25
136.2 C 222kPa 4:55
136.4 C 225kPa 5:25
136.2 C 223kPa 5:55
136.3 C 223kPa 6:25
136.4 C 223kPa 6:55
136.4 C 221kPa 7:25
136.4 C 221kPa 7:25
VENTING 7:25
AIR DRYING 8:18
CYCLE COMPLETE 8:32

Digital Signature #
675F560794114359
