

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4458

Name: RITTER 1  
Sterilizer ID: V2178198  
04/15/2024 12:47 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.2 F 0.0  
2:00 151.4 F 2.4  
4:00 165.1 F 5.3  
6:00 193.8 F 9.3  
8:00 212.0 F 11.1  
10:00 220.5 F 6.1  
12:00 231.9 F 10.2  
14:00 234.0 F 8.8  
16:00 241.4 F 11.6  
18:00 238.3 F 9.6  
20:00 250.6 F 15.6  
22:00 261.3 F 21.6  
24:00 271.0 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.7  
1:00 271.8 F 28.7  
2:00 272.8 F 29.0  
3:00 272.4 F 29.3  
4:00 271.5 F 28.8

Min 271.4 F 28.6  
Max 272.9 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.6 F 0.0  
5:00 226.8 F 0.0  
10:00 208.4 F 0.0  
15:00 198.1 F 0.0  
20:00 203.7 F 0.0  
25:00 189.5 F 0.0  
30:00 176.2 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 25:25  
STERILIZING: 4:00  
VENTING: 3:49  
DRYING: 30:00  
TOTAL CYCLE: 01:04:19  
CYCLE COMPLETE

04/15/2024 01:51 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_