
Midmark M11 - v1.0.0
Total Cycles: 1954

Name: RITTER 1
Sterilizer ID: V2178198
05/26/2022 07:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.7 F 0.0
2:00 75.0 F 0.0
4:00 105.0 F 0.0
6:00 122.2 F 1.0
8:00 144.3 F 3.7
10:00 176.0 F 7.2
12:00 225.9 F 11.1
14:00 220.9 F 4.5
16:00 228.6 F 7.7
18:00 237.6 F 11.5
20:00 228.5 F 6.5
22:00 242.2 F 11.9
24:00 237.4 F 9.1
26:00 250.3 F 15.2
28:00 260.9 F 21.2
30:00 270.7 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 271.6 F 28.7
2:00 272.1 F 28.8
3:00 272.8 F 29.0
4:00 272.3 F 29.2

Min 271.2 F 28.6
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.8 F 0.0
5:00 222.6 F 0.0
10:00 195.8 F 0.0
15:00 220.7 F 0.0
20:00 221.9 F 0.0
25:00 219.9 F 0.0
30:00 202.0 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 31:30
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:09:29
CYCLE COMPLETE

05/26/2022 08:37 AM
ACCEPTED: X
REJECTED:
Operator: U02