
Midmark M11 - v1.0.0
Total Cycles: 3185

Name: RITTER 1
Sterilizer ID: V2178198
05/02/2023 08:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 189.3 F 0.0
2:00 177.6 F 0.0
4:00 185.9 F 2.5
6:00 216.2 F 7.0
8:00 230.4 F 12.0
10:00 228.7 F 7.5
12:00 240.6 F 12.7
14:00 238.2 F 10.3
16:00 236.8 F 8.5
18:00 250.0 F 15.5
20:00 262.1 F 22.4
22:00 272.7 F 29.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.7 F 29.4
2:00 271.8 F 28.7
3:00 272.1 F 29.2
4:00 272.5 F 29.3

Min 271.5 F 28.5
Max 272.7 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.3 F 0.0
5:00 232.7 F 0.0
10:00 224.8 F 0.0
15:00 231.2 F 0.0
20:00 227.0 F 0.0
25:00 229.2 F 0.0
30:00 228.7 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 22:23
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:51
CYCLE COMPLETE

05/02/2023 09:37 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____