

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2623

Name: RITTER2  
Sterilizer ID: V2313228  
07/07/2023 11:04 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 210.3 F 0.0  
2:00 206.5 F 4.8  
4:00 217.4 F 12.9  
6:00 231.4 F 10.1  
8:00 236.4 F 9.7  
10:00 237.9 F 9.8  
12:00 255.5 F 18.4

C103: HEATUP MODE  
STOP PRESSED  
CHAMBER TEMP 252F  
CHAMBER PRESS 16.2PSI  
\*\* ITEMS NOT STERILE \*\*

07/07/2023 11:19 AM