
Midmark M11 - v1.0.5
Total Cycles: 4013

Name: RITTER2
Sterilizer ID: V2313228
10/09/2024 11:39 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 105.5 F 0.0
2:00 104.4 F 0.0
4:00 109.3 F 0.0
6:00 140.4 F 0.2
8:00 178.8 F 2.4
10:00 199.0 F 5.7
12:00 211.7 F 9.1
14:00 225.0 F 12.8
16:00 226.9 F 7.5
18:00 236.2 F 12.0
20:00 234.4 F 9.2
22:00 239.8 F 10.7
24:00 245.0 F 13.4
26:00 255.9 F 19.3
28:00 266.2 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 29.2
1:00 272.4 F 30.0
2:00 271.8 F 29.3
3:00 271.6 F 29.7
4:00 272.4 F 29.8

Min 271.2 F 29.1
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.6 F 0.0
5:00 206.5 F 0.0
10:00 219.7 F 0.0
15:00 219.4 F 0.0
20:00 225.7 F 0.0
25:00 234.4 F 0.0
30:00 226.3 F 0.0
Drying Complete

FILLING: 0:52
HEATING: 29:49
STERILIZING: 4:00
VENTING: 3:49
DRYING: 30:00
TOTAL CYCLE: 01:08:29
CYCLE COMPLETE

10/09/2024 12:47 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____