
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002605
10:48 25/03/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:16
PRESSURIZING 3:17
STERILIZING 4:22
136.0 C 222kPa 4:22
136.3 C 225kPa 4:52
136.2 C 223kPa 5:22
136.3 C 225kPa 5:52
136.3 C 225kPa 6:22
136.4 C 225kPa 6:52
136.4 C 225kPa 7:22
136.3 C 224kPa 7:52
136.3 C 224kPa 7:52
VENTING 7:52
AIR DRYING 8:51
CYCLE COMPLETE 8:55

Digital Signature #
138EDAC95F7FA486
