
Midmark M11 - v1.0.0
Total Cycles: 5014

Name: RITTER 1
Sterilizer ID: V2178198
10/22/2024 10:18 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 144.7 F 0.0
2:00 141.4 F 0.1
4:00 144.3 F 0.4
6:00 174.0 F 0.9
8:00 189.6 F 4.3
10:00 204.4 F 8.0
12:00 217.2 F 11.9
14:00 223.5 F 6.1
16:00 233.6 F 10.8
18:00 233.6 F 7.6
20:00 244.6 F 13.1
22:00 242.5 F 11.4
24:00 254.2 F 17.3
26:00 264.5 F 23.4
28:00 271.8 F 28.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.1
1:00 272.4 F 28.8
2:00 271.7 F 28.2
3:00 271.6 F 28.5
4:00 272.2 F 28.7

Min 271.2 F 28.0
Max 272.4 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 198.7 F 0.0
5:00 193.2 F 0.0
10:00 202.7 F 0.0
15:00 214.8 F 0.0
20:00 226.1 F 0.0
25:00 231.9 F 0.0
30:00 229.4 F 0.0
Drying Complete

FILLING: 0:46
HEATING: 28:10
STERILIZING: 4:00
VENTING: 3:42
DRYING: 30:00
TOTAL CYCLE: 01:06:38
CYCLE COMPLETE

10/22/2024 11:24 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____