

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 002135  
7:14 27/07/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:19  
PRESSURIZING 2:23  
STERILIZING 3:26  
136.4 C 220kPa 3:26  
136.4 C 223kPa 3:56  
136.4 C 222kPa 4:26  
136.4 C 222kPa 4:56  
136.4 C 224kPa 5:26  
136.4 C 222kPa 5:56  
136.2 C 222kPa 6:26  
136.2 C 220kPa 6:56  
136.2 C 220kPa 6:56  
VENTING 6:56  
AIR DRYING 7:48  
CYCLE COMPLETE 7:59

-----  
Digital Signature #  
F06595A3BE5B578F  
-----