
Midmark M11 - v1.0.0
Total Cycles: 4543

Name: RITTER 1
Sterilizer ID: V2178198
05/10/2024 01:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.1 F 0.0
2:00 105.4 F 0.6
4:00 122.4 F 2.7
6:00 150.0 F 5.5
8:00 174.4 F 8.8
10:00 192.9 F 12.5
12:00 213.9 F 4.6
14:00 222.2 F 8.7
16:00 232.5 F 13.0
18:00 235.9 F 9.3
20:00 239.8 F 10.3
22:00 244.8 F 12.6
24:00 256.0 F 18.5
26:00 266.4 F 24.9
28:00 273.0 F 29.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.7
1:00 271.6 F 28.7
2:00 272.6 F 29.0
3:00 272.4 F 29.2
4:00 271.4 F 28.9

Min 271.2 F 28.6
Max 272.9 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 190.5 F 0.0
5:00 195.4 F 0.0
10:00 189.0 F 0.0
15:00 187.0 F 0.0
20:00 184.9 F 0.0
25:00 183.7 F 0.0
30:00 182.4 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 28:15
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:06:44
CYCLE COMPLETE

05/10/2024 02:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____