
Midmark M11 - v1.0.5
Total Cycles: 2813

Name: RITTER2
Sterilizer ID: V2313228
08/29/2023 03:00 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 175.5 F 0.0
2:00 176.0 F 1.0
4:00 203.0 F 5.0
6:00 219.0 F 9.5
8:00 230.3 F 9.8
10:00 234.8 F 10.6
12:00 236.7 F 8.3
14:00 240.9 F 10.8
16:00 247.4 F 14.3
18:00 259.3 F 20.8
20:00 269.9 F 27.7

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.8
1:00 272.6 F 29.5
2:00 271.2 F 28.7
3:00 272.3 F 29.7
4:00 271.8 F 29.0

Min 271.1 F 28.7
Max 272.7 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 212.0 F 0.0
10:00 215.0 F 0.0
15:00 216.6 F 0.0
20:00 215.0 F 0.0
25:00 215.1 F 0.0
30:00 214.1 F 0.0
Drying Complete

FILLING: 2:16
HEATING: 20:57
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:00:44
CYCLE COMPLETE

08/29/2023 04:00 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____