

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2335

Name: RITTER2  
Sterilizer ID: V2313228  
05/03/2023 04:14 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 201.4 F 0.0  
2:00 201.9 F 4.0  
4:00 208.2 F 9.4  
6:00 222.7 F 7.9  
8:00 233.0 F 10.6  
10:00 234.9 F 8.2  
12:00 240.0 F 10.7  
14:00 247.6 F 14.6  
16:00 260.2 F 21.6  
18:00 271.4 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 272.1 F 29.1  
1:00 272.5 F 29.8  
2:00 271.6 F 29.0  
3:00 272.5 F 29.7  
4:00 271.6 F 29.0  
  
Min 271.5 F 29.0  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 206.9 F 0.0  
5:00 210.4 F 0.0  
10:00 224.9 F 0.0  
15:00 218.5 F 0.0  
20:00 217.5 F 0.0  
25:00 221.7 F 0.0  
30:00 218.0 F 0.0  
  
Drying Complete

FILLING: 2:22  
HEATING: 18:29  
STERILIZING: 4:00  
VENTING: 3:27  
DRYING: 30:00  
TOTAL CYCLE: 00:58:18  
CYCLE COMPLETE

05/03/2023 05:12 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_