

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004909  
10:27 05/01/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:53  
PRESSURIZING 1:57  
STERILIZING 2:58  
136.4 C 221kPa 2:58  
136.4 C 224kPa 3:28  
136.2 C 222kPa 3:58  
136.4 C 224kPa 4:28  
136.2 C 217kPa 4:58  
136.2 C 221kPa 5:28  
136.3 C 223kPa 5:58  
136.4 C 224kPa 6:28  
136.4 C 224kPa 6:28  
VENTING 6:28  
AIR DRYING 7:18  
CYCLE COMPLETE 12:10

-----  
Digital Signature #  
29A1AB570C3E2011  
-----