
Midmark M11 - v1.0.0
Total Cycles: 1136

Name: RITTER 1
Sterilizer ID: V2178198
08/31/2021 12:11 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 148.0 F 0.0
2:00 149.5 F 1.1
4:00 167.4 F 3.4
6:00 187.6 F 6.3
8:00 203.8 F 10.2
10:00 220.8 F 8.2
12:00 223.9 F 7.8
14:00 235.6 F 12.3
16:00 234.1 F 9.0
18:00 242.3 F 11.2
20:00 244.7 F 12.8
22:00 256.0 F 18.8
24:00 266.4 F 25.1

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.6
1:00 272.4 F 29.3
2:00 272.6 F 29.1
3:00 271.7 F 28.7
4:00 271.2 F 28.5

Min 271.2 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.4 F 0.0
5:00 208.1 F 0.0
10:00 220.7 F 0.0
15:00 217.0 F 0.0
20:00 208.4 F 0.0
25:00 212.3 F 0.0
30:00 211.4 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 25:47
STERILIZING: 4:00
VENTING: 3:31
DRYING: 30:00
TOTAL CYCLE: 01:04:28
CYCLE COMPLETE

08/31/2021 01:15 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____