
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000048
12:59 28/02/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:18
PRESSURIZING 2:24
STERILIZING 3:31
136.0 C 221kPa 3:31
136.0 C 222kPa 4:01
136.4 C 222kPa 4:31
136.4 C 225kPa 5:01
136.4 C 223kPa 5:31
136.4 C 223kPa 6:01
136.1 C 223kPa 6:31
136.2 C 224kPa 7:01
136.2 C 224kPa 7:01
VENTING 7:01
AIR DRYING 7:52
CYCLE COMPLETE 8:06

Digital Signature #
8B8433685F251C9C
