

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4215

Name: RITTER2  
Sterilizer ID: V2313228  
12/13/2024 05:56 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.0 F 0.0  
2:00 180.2 F 3.6  
4:00 206.6 F 9.6

C103: HEATUP MODE  
STOP PRESSED  
CHAMBER TEMP 217F  
CHAMBER PRESS 12.8PSI  
\*\* ITEMS NOT STERILE \*\*

12/13/2024 06:03 PM