
Midmark M11 - v1.0.0
Total Cycles: 2624

Name: RITTER 1
Sterilizer ID: V2178198
11/29/2022 06:43 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.1 F 0.0
2:00 199.4 F 5.2
4:00 219.3 F 11.1
6:00 222.5 F 6.5
8:00 239.3 F 12.8
10:00 238.0 F 10.5
12:00 237.7 F 9.2
14:00 252.9 F 16.7
16:00 265.7 F 24.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.3 F 29.0
2:00 271.5 F 28.3
3:00 271.7 F 28.9
4:00 272.2 F 29.0

Min 271.1 F 28.2
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 215.6 F 0.0
5:00 225.5 F 0.0
10:00 240.6 F 0.0
15:00 235.1 F 0.0
20:00 226.7 F 0.0
25:00 227.2 F 0.0
30:00 226.7 F 0.0
Drying Complete

FILLING: 1:48
HEATING: 17:37
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 00:56:31
CYCLE COMPLETE

11/29/2022 07:39 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____