

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1734

Name: RITTER2  
Sterilizer ID: V2313228  
11/16/2022 01:41 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 176.0 F 0.0  
2:00 173.2 F 0.4  
4:00 203.3 F 2.8  
6:00 220.7 F 7.9  
8:00 235.2 F 13.2  
10:00 236.7 F 10.2  
12:00 234.7 F 8.5  
14:00 240.7 F 10.6  
16:00 248.3 F 14.7  
18:00 260.8 F 21.8  
20:00 272.2 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.9  
1:00 272.5 F 29.4  
2:00 271.7 F 29.1  
3:00 272.5 F 29.4  
4:00 271.5 F 28.9

Min 271.5 F 28.8  
Max 272.7 F 29.7

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 183.6 F 0.0  
5:00 227.4 F 0.0  
10:00 234.6 F 0.0  
15:00 232.1 F 0.0  
20:00 226.4 F 0.0  
25:00 231.1 F 0.0  
30:00 223.0 F 0.0  
Drying Complete

FILLING: 1:12  
HEATING: 20:29  
STERILIZING: 4:00  
VENTING: 3:12  
DRYING: 30:00  
TOTAL CYCLE: 00:58:53  
CYCLE COMPLETE

11/16/2022 02:39 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_