
Midmark M11 - v1.0.5
Total Cycles: 4040

Name: RITTER2
Sterilizer ID: V2313228
10/17/2024 03:04 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 156.8 F 0.0
2:00 163.2 F 3.1
4:00 190.7 F 7.8
6:00 212.1 F 12.1
8:00 225.4 F 8.1
10:00 240.4 F 12.6
12:00 240.6 F 11.6
14:00 240.5 F 10.7
16:00 254.6 F 17.9
18:00 266.6 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.7
1:00 272.3 F 28.9
2:00 272.8 F 29.4
3:00 272.0 F 29.0
4:00 272.4 F 29.0

Min 271.8 F 28.6
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 184.1 F 0.0
5:00 193.6 F 0.0
10:00 204.6 F 0.0
15:00 206.7 F 0.0
20:00 205.3 F 0.0
25:00 205.8 F 0.0
30:00 204.9 F 0.0

Drying Complete

FILLING: 1:43
HEATING: 19:26
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:25
CYCLE COMPLETE

10/17/2024 04:02 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____