

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2782

Name: RITTER 1  
Sterilizer ID: V2178198  
01/17/2023 09:24 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 75.0 F 0.0  
2:00 75.7 F 0.0  
4:00 82.5 F 0.0  
6:00 113.3 F 0.8  
8:00 142.3 F 2.3  
10:00 167.6 F 4.4  
12:00 189.6 F 7.5  
14:00 208.2 F 11.2  
16:00 215.2 F 4.1  
18:00 224.6 F 7.5  
20:00 235.2 F 11.7  
22:00 233.8 F 8.3  
24:00 242.9 F 11.7  
26:00 244.6 F 12.6  
28:00 256.6 F 18.9  
30:00 267.5 F 25.6

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.3  
1:00 272.5 F 29.2  
2:00 271.9 F 28.6  
3:00 271.2 F 28.4  
4:00 272.6 F 29.3

Min 271.1 F 28.2  
Max 272.8 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.0 F 0.0  
5:00 230.2 F 0.0  
10:00 224.3 F 0.0  
15:00 227.7 F 0.0  
20:00 228.8 F 0.0  
25:00 228.2 F 0.0  
30:00 229.8 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 31:29  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 01:09:26  
CYCLE COMPLETE

01/17/2023 10:33 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_