
Midmark M11 - v1.0.0
Total Cycles: 3914

Name: RITTER 1
Sterilizer ID: V2178198
10/17/2023 03:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 192.0 F 0.0
2:00 187.2 F 4.1
4:00 204.3 F 9.0
6:00 222.6 F 7.0
8:00 231.9 F 9.9
10:00 234.3 F 8.1
12:00 244.3 F 12.3
14:00 245.3 F 13.0
16:00 258.4 F 20.1
18:00 270.2 F 27.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.5
1:00 272.3 F 28.9
2:00 272.0 F 29.0
3:00 272.1 F 28.8
4:00 271.8 F 28.9

Min 271.5 F 28.4
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.5 F 0.0
5:00 228.6 F 0.0
10:00 231.7 F 0.0
15:00 235.8 F 0.0
20:00 226.6 F 0.0
25:00 217.0 F 0.0
30:00 215.4 F 0.0
Drying Complete

FILLING: 2:16
HEATING: 18:48
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

10/17/2023 04:00 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____