
Mississauga, ON
Canada

STATIM 5000 S501R751
SN 511020J00004
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000820
14:11 28/03/2022
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:36
PRESSURIZING 6:24
STERILIZING 7:56
136.2 C 220kPa 7:56
135.7 C 218kPa 8:26
136.4 C 223kPa 8:56
136.4 C 223kPa 9:26
136.4 C 223kPa 9:56
136.4 C 221kPa 10:26
136.4 C 222kPa 10:56
136.4 C 223kPa 11:26
136.4 C 223kPa 11:26
VENTING 11:26
AIR DRYING 12:30
CYCLE COMPLETE 25:17

Digital Signature #
C445CEB31FCC23C1
