
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000141
14:03 11/09/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:36
PRESSURIZING 3:38
STERILIZING 4:45
136.0 C 222kPa 4:45
136.5 C 226kPa 5:15
136.4 C 226kPa 5:45
136.4 C 224kPa 6:15
136.4 C 225kPa 6:45
136.3 C 224kPa 7:15
135.9 C 222kPa 7:45
136.4 C 225kPa 8:15
136.4 C 225kPa 8:15
VENTING 8:15
AIR DRYING 9:10
CYCLE COMPLETE 24:10

Digital Signature #
E5F1132489AD9C19
