
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R750
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 001026
12:36 21/12/2020
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:33
PRESSURIZING 2:36
STERILIZING 3:37
136.0 C 221kPa 3:37
136.2 C 223kPa 4:07
136.2 C 221kPa 4:37
136.3 C 223kPa 5:07
136.4 C 222kPa 5:37
136.4 C 224kPa 6:07
136.3 C 223kPa 6:37
136.3 C 222kPa 7:07
136.3 C 222kPa 7:07
VENTING 7:07
AIR DRYING 7:57
CYCLE COMPLETE 8:11

Digital Signature #
E4AD19723DB259D6
