

-----

Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005261  
10:05 24/03/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:51  
PRESSURIZING 1:55  
STERILIZING 2:43  
136.0 C 220kPa 2:43  
136.4 C 223kPa 3:13  
136.3 C 221kPa 3:43  
136.1 C 220kPa 4:13  
136.4 C 224kPa 4:43  
136.4 C 224kPa 5:13  
136.4 C 224kPa 5:43  
136.3 C 223kPa 6:13  
136.3 C 223kPa 6:13  
VENTING 6:13  
AIR DRYING 7:02  
CYCLE COMPLETE 14:10  
-----

Digital Signature #  
EE107EBCA9D3638E  
-----