
Midmark M11 - v1.0.5
Total Cycles: 2686

Name: RITTER2
Sterilizer ID: V2313228
07/24/2023 10:53 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 169.6 F 0.0
2:00 170.1 F 1.0
4:00 175.4 F 4.2
6:00 195.4 F 8.1
8:00 210.1 F 12.1
10:00 220.0 F 6.1
12:00 230.5 F 10.4
14:00 233.9 F 8.9
16:00 242.1 F 12.7
18:00 241.6 F 11.5
20:00 253.2 F 17.5
22:00 263.5 F 23.6
24:00 272.7 F 29.6

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.0
1:00 272.3 F 29.8
2:00 272.2 F 29.3
3:00 271.2 F 28.9
4:00 271.6 F 29.6

Min 271.0 F 28.8
Max 272.7 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.8 F 0.0
5:00 198.4 F 0.0
10:00 201.8 F 0.0
15:00 198.9 F 0.0
20:00 195.3 F 0.0
25:00 195.5 F 0.0
30:00 195.3 F 0.0
Drying Complete

FILLING: 1:41
HEATING: 24:21
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:03:43
CYCLE COMPLETE

07/24/2023 11:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____