
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 000777
7:12 29/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:18
PRESSURIZING 1:26
STERILIZING 2:35
136.4 C 231kPa 2:35
136.2 C 231kPa 3:05
136.4 C 233kPa 3:35
136.2 C 230kPa 4:05
136.1 C 230kPa 4:35
136.4 C 233kPa 5:05
136.3 C 231kPa 5:35
136.3 C 231kPa 6:05
136.3 C 231kPa 6:05
VENTING 6:05
AIR DRYING 6:57
CYCLE COMPLETE 8:24

Digital Signature #
E61FE5601D7FB563
