

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4861

Name: RITTER 1  
Sterilizer ID: V2178198  
08/29/2024 08:29 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 70.3 F 0.0  
2:00 74.0 F 0.0  
4:00 86.8 F 0.0  
6:00 110.8 F 0.3  
8:00 144.4 F 1.0  
10:00 174.9 F 1.1  
12:00 198.3 F 4.1  
14:00 213.0 F 7.4  
16:00 222.7 F 11.0  
18:00 224.1 F 5.9  
20:00 235.2 F 9.8  
22:00 237.0 F 9.0  
24:00 244.2 F 12.3  
26:00 241.3 F 10.7  
28:00 254.1 F 17.1  
30:00 265.2 F 23.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.7 F 28.0  
1:00 272.4 F 28.3  
2:00 272.4 F 28.7  
3:00 271.8 F 27.9  
4:00 272.7 F 28.7

Min 271.6 F 27.8  
Max 272.8 F 28.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.8 F 0.0  
5:00 194.6 F 0.0  
10:00 192.6 F 0.0  
15:00 231.8 F 0.0  
20:00 230.4 F 0.0  
25:00 229.5 F 0.0  
30:00 229.4 F 0.0  
Drying Complete

FILLING: 0:41  
HEATING: 31:49  
STERILIZING: 4:00  
VENTING: 3:18  
DRYING: 30:00  
TOTAL CYCLE: 01:09:48  
CYCLE COMPLETE

08/29/2024 09:38 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_