

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2222

Name: RITTER 1  
Sterilizer ID: V2178198  
08/15/2022 11:45 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.2 F 0.0  
2:00 194.6 F 4.5  
4:00 203.1 F 9.6  
6:00 218.1 F 6.7  
8:00 232.0 F 9.8  
10:00 235.4 F 8.6  
12:00 243.5 F 12.7  
14:00 241.4 F 11.2  
16:00 254.6 F 17.8  
18:00 266.1 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.5  
1:00 272.3 F 28.8  
2:00 272.2 F 29.1  
3:00 271.8 F 28.5  
4:00 272.5 F 29.1  
  
Min 271.5 F 28.3  
Max 272.5 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 212.2 F 0.0  
5:00 230.9 F 0.0  
10:00 231.3 F 0.0  
15:00 234.6 F 0.0  
20:00 226.4 F 0.0  
25:00 219.1 F 0.0  
30:00 222.2 F 0.0  
Drying Complete

FILLING: 1:40  
HEATING: 19:37  
STERILIZING: 4:00  
VENTING: 3:11  
DRYING: 30:00  
TOTAL CYCLE: 00:58:28  
CYCLE COMPLETE

08/15/2022 12:43 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_