
Midmark M11 - v1.0.0
Total Cycles: 1123

Name: RITTER 1
Sterilizer ID: V2178198
08/27/2021 08:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 145.8 F 0.0
2:00 145.9 F 0.8
4:00 151.8 F 2.4
6:00 180.9 F 5.3
8:00 200.3 F 9.3
10:00 218.9 F 10.4
12:00 223.8 F 7.0
14:00 236.2 F 11.2
16:00 235.5 F 7.5
18:00 243.8 F 13.1
20:00 242.4 F 11.5
22:00 254.2 F 17.6
24:00 264.8 F 23.9
26:00 271.3 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.4
1:00 272.6 F 29.1
2:00 272.3 F 28.7
3:00 271.3 F 28.3
4:00 271.9 F 28.9

Min 271.3 F 28.2
Max 272.8 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 218.0 F 0.0
10:00 226.3 F 0.0
15:00 233.8 F 0.0
20:00 245.5 F 0.0
25:00 231.6 F 0.0
30:00 227.9 F 0.0
Drying Complete

FILLING: 1:11
HEATING: 26:04
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 01:04:38
CYCLE COMPLETE

08/27/2021 09:58 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____