
Midmark M11 - v1.0.0
Total Cycles: 1234

Name: RITTER 1
Sterilizer ID: V2178198
10/06/2021 09:10 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 95.4 F 0.0
2:00 98.2 F 0.0
4:00 103.4 F 0.0
6:00 134.1 F 0.2
8:00 172.4 F 0.3
10:00 198.2 F 3.6
12:00 213.0 F 7.3
14:00 223.0 F 11.2
16:00 225.8 F 5.9
18:00 237.2 F 10.8
20:00 234.2 F 7.7
22:00 245.6 F 12.4
24:00 245.5 F 12.6
26:00 257.5 F 19.0
28:00 268.4 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.8 F 28.6
2:00 271.5 F 27.8
3:00 272.7 F 28.7
4:00 271.9 F 28.1

Min 271.4 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 195.8 F 0.0
5:00 213.7 F 0.0
10:00 205.9 F 0.0
15:00 217.8 F 0.0
20:00 189.6 F 0.0
25:00 170.3 F 0.0
30:00 166.5 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 29:15
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 01:07:32
CYCLE COMPLETE

10/06/2021 10:20 AM
ACCEPTED: X
REJECTED:
Operator: U01