
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 003857
9:36 22/06/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:13
PRESSURIZING 2:18
STERILIZING 3:20
136.0 C 220kPa 3:20
136.4 C 225kPa 3:50
136.4 C 223kPa 4:20
136.4 C 223kPa 4:50
136.4 C 223kPa 5:20
136.0 C 222kPa 5:50
135.9 C 220kPa 6:20
136.4 C 225kPa 6:50
136.4 C 225kPa 6:50
VENTING 6:50
AIR DRYING 8:41
CYCLE COMPLETE 9:45

Digital Signature #
FFC1C262266F969E
