
Midmark M11 - v1.0.5
Total Cycles: 209

Name: RITTER2
Sterilizer ID: V2313228
07/14/2021 01:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 201.6 F 0.0
2:00 198.0 F 4.1
4:00 210.5 F 9.9
6:00 224.7 F 7.3
8:00 236.4 F 12.1
10:00 237.3 F 10.4
12:00 236.9 F 9.5
14:00 251.4 F 16.5
16:00 263.1 F 23.4
18:00 271.8 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.0
1:00 271.9 F 29.0
2:00 272.3 F 29.3
3:00 272.6 F 29.7
4:00 271.7 F 29.2

Min 271.6 F 28.9
Max 272.6 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 217.7 F 0.0
10:00 226.7 F 0.0
15:00 230.4 F 0.0
20:00 226.0 F 0.0
25:00 217.0 F 0.0
30:00 254.1 F 0.0
Drying Complete

FILLING: 2:49
HEATING: 18:03
STERILIZING: 4:00
VENTING: 3:13
DRYING: 30:00
TOTAL CYCLE: 00:58:05
CYCLE COMPLETE

07/14/2021 02:31 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____