

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2110

Name: RITTER2  
Sterilizer ID: V2313228  
03/01/2023 02:40 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 129.9 F 0.0  
2:00 134.1 F 1.8  
4:00 153.4 F 5.0  
6:00 177.8 F 9.4  
8:00 209.6 F 6.9  
10:00 221.0 F 8.7  
12:00 230.3 F 10.7  
14:00 236.7 F 11.2  
16:00 236.5 F 9.6  
18:00 250.3 F 16.2  
20:00 262.3 F 23.1  
22:00 272.6 F 29.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.4  
1:00 272.4 F 29.8  
2:00 271.5 F 29.5  
3:00 272.4 F 29.9  
4:00 271.6 F 29.2

Min 271.4 F 29.2  
Max 272.5 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 197.3 F 0.0  
5:00 222.3 F 0.0  
10:00 229.5 F 0.0  
15:00 222.9 F 0.0  
20:00 230.7 F 0.0  
25:00 231.6 F 0.0  
30:00 225.6 F 0.0  
Drying Complete

FILLING: 1:04  
HEATING: 22:17  
STERILIZING: 4:00  
VENTING: 3:15  
DRYING: 30:00  
TOTAL CYCLE: 01:00:36  
CYCLE COMPLETE

03/01/2023 03:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_