
Midmark M11 - v1.0.0
Total Cycles: 1188

Name: RITTER 1
Sterilizer ID: V2178198
09/21/2021 06:37 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.3 F 0.0
2:00 74.2 F 0.0
4:00 83.3 F 0.0
6:00 107.6 F 0.1
8:00 141.5 F 0.1
10:00 174.0 F 0.2
12:00 195.8 F 0.3
14:00 211.5 F 3.3
16:00 220.8 F 5.9
18:00 229.2 F 8.8
20:00 237.6 F 12.3
22:00 233.6 F 8.7
24:00 244.0 F 12.4
26:00 243.3 F 12.0
28:00 240.9 F 10.7
30:00 253.4 F 16.8
32:00 264.1 F 23.1
34:00 272.0 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.7 F 28.8
2:00 271.9 F 28.1
3:00 272.0 F 28.4
4:00 272.7 F 28.6

Min 271.5 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.7 F 0.0
5:00 182.1 F 0.0
10:00 193.8 F 0.0
15:00 192.1 F 0.0
20:00 200.0 F 0.0
25:00 196.2 F 0.0
30:00 209.2 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 34:11
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:12:27
CYCLE COMPLETE

09/21/2021 07:56 AM
ACCEPTED: X
REJECTED:
Operator: U01