

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000674  
13:20 14/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:34  
PRESSURIZING 2:40  
STERILIZING 3:42  
136.0 C 228kPa 3:42  
136.4 C 232kPa 4:12  
136.4 C 233kPa 4:42  
136.3 C 232kPa 5:12  
136.4 C 233kPa 5:42  
136.4 C 233kPa 6:12  
136.4 C 233kPa 6:42  
136.4 C 231kPa 7:12  
136.4 C 231kPa 7:12  
VENTING 7:12  
AIR DRYING 8:10  
CYCLE COMPLETE 8:17  
-----

Digital Signature #  
EAF8BFE7C1B0EC01  
-----