

-----  
Midmark M11 - v1.0.0  
Total Cycles: 1651

Name: RITTER 1  
Sterilizer ID: V2178198  
02/17/2022 03:00 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 186.4 F 0.0  
2:00 177.2 F 2.9  
4:00 190.9 F 7.1  
6:00 233.5 F 12.4  
8:00 231.5 F 7.9  
10:00 241.2 F 13.1  
12:00 239.6 F 10.6  
14:00 235.8 F 8.6  
16:00 250.6 F 15.6  
18:00 262.5 F 22.5  
20:00 272.6 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.3 F 28.7  
1:00 272.6 F 28.9  
2:00 272.5 F 29.3  
3:00 271.7 F 28.6  
4:00 272.5 F 29.4  
  
Min 271.3 F 28.5  
Max 272.9 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.9 F 0.0  
5:00 231.3 F 0.0  
10:00 231.6 F 0.0  
15:00 231.3 F 0.0  
20:00 221.8 F 0.0  
25:00 226.3 F 0.0  
30:00 229.0 F 0.0  
Drying Complete

FILLING: 2:14  
HEATING: 20:15  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:59:37  
CYCLE COMPLETE

02/17/2022 03:59 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_