
Midmark M11 - v1.0.5
Total Cycles: 3395

Name: RITTER2
Sterilizer ID: V2313228
02/21/2024 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 150.5 F 0.0
2:00 154.5 F 1.4
4:00 164.7 F 5.1
6:00 191.7 F 9.8
8:00 211.7 F 7.7
10:00 225.4 F 10.2
12:00 232.2 F 7.8
14:00 239.6 F 12.8
16:00 241.0 F 11.7
18:00 254.0 F 18.3
20:00 265.5 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.3
1:00 273.2 F 30.0
2:00 271.3 F 29.2
3:00 272.2 F 30.1
4:00 271.9 F 29.8

Min 270.9 F 29.2
Max 273.5 F 30.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.3 F 0.0
5:00 218.4 F 0.0
10:00 230.2 F 0.0
15:00 213.0 F 0.0
20:00 206.0 F 0.0
25:00 205.6 F 0.0
30:00 205.4 F 0.0
Drying Complete

FILLING: 1:19
HEATING: 21:45
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:00:28
CYCLE COMPLETE

02/21/2024 12:01 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____