
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000962
15:31 07/03/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:20
PRESSURIZING 3:26
STERILIZING 4:31
136.0 C 228kPa 4:31
136.4 C 233kPa 5:01
136.4 C 233kPa 5:31
136.4 C 231kPa 6:01
136.4 C 233kPa 6:31
136.4 C 231kPa 7:01
136.3 C 232kPa 7:31
136.2 C 231kPa 8:01
136.2 C 231kPa 8:01
VENTING 8:01
AIR DRYING 8:54
CYCLE COMPLETE 9:04

Digital Signature #
4DAD5496FB3D66D8
