

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4346

Name: RITTER2  
Sterilizer ID: V2313228  
02/10/2025 02:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.9 F 0.0  
2:00 145.8 F 2.1  
4:00 155.4 F 5.3  
6:00 180.4 F 9.2  
8:00 198.1 F 13.3  
10:00 216.5 F 5.8  
12:00 226.9 F 10.0  
14:00 233.1 F 9.7  
16:00 239.2 F 11.6  
18:00 237.5 F 9.8  
20:00 249.7 F 15.7  
22:00 260.3 F 21.7  
24:00 270.0 F 28.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.2  
1:00 272.0 F 29.9  
2:00 272.4 F 29.8  
3:00 272.1 F 29.5  
4:00 271.4 F 29.1

Min 271.1 F 29.0  
Max 272.6 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.4 F 0.0  
5:00 198.8 F 0.0  
10:00 178.1 F 0.0  
15:00 177.2 F 0.0  
20:00 170.8 F 0.0  
25:00 164.5 F 0.0  
30:00 170.2 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 25:02  
STERILIZING: 4:00  
VENTING: 3:54  
DRYING: 30:00  
TOTAL CYCLE: 01:04:03  
CYCLE COMPLETE

02/10/2025 03:59 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_