
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 007487
9:24 04/06/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:10
STERILIZING 4:14
136.0 C 220kPa 4:14
136.4 C 225kPa 4:44
136.1 C 219kPa 5:14
136.4 C 226kPa 5:44
136.4 C 224kPa 6:14
136.3 C 220kPa 6:44
135.9 C 220kPa 7:14
136.4 C 225kPa 7:44
136.4 C 225kPa 7:44
VENTING 7:44
AIR DRYING 8:44
CYCLE COMPLETE 9:26

Digital Signature #
C4DC8416351D0D90
