
Midmark M11 - v1.0.0
Total Cycles: 4269

Name: RITTER 1
Sterilizer ID: V2178198
02/05/2024 11:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 155.5 F 0.0
2:00 154.1 F 2.6
4:00 164.4 F 5.9
6:00 191.1 F 9.4
8:00 222.3 F 8.2
10:00 220.7 F 6.6
12:00 232.4 F 10.8
14:00 234.9 F 7.3
16:00 242.7 F 12.4
18:00 240.3 F 10.6
20:00 252.6 F 16.6
22:00 263.1 F 22.8
24:00 272.6 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.3
1:00 272.7 F 29.0
2:00 271.3 F 28.2
3:00 272.2 F 29.1
4:00 272.0 F 28.6

Min 271.0 F 28.2
Max 272.7 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.6 F 0.0
5:00 220.7 F 0.0
10:00 229.0 F 0.0
15:00 230.7 F 0.0
20:00 232.9 F 0.0
25:00 228.2 F 0.0
30:00 224.2 F 0.0
Drying Complete

FILLING: 1:06
HEATING: 24:20
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:03:12
CYCLE COMPLETE

02/05/2024 12:54 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____