
Midmark M11 - v1.0.5
Total Cycles: 3269

Name: RITTER2
Sterilizer ID: V2313228
01/03/2024 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 141.8 F 0.0
2:00 141.5 F 0.7
4:00 143.4 F 2.8
6:00 167.7 F 6.5
8:00 192.2 F 11.1
10:00 212.2 F 4.4
12:00 226.2 F 9.3
14:00 233.2 F 10.1
16:00 240.2 F 11.7
18:00 238.4 F 10.1
20:00 251.3 F 16.5
22:00 262.6 F 23.1
24:00 272.9 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 29.0
1:00 272.6 F 29.8
2:00 271.6 F 29.0
3:00 271.9 F 29.7
4:00 272.2 F 29.5

Min 271.1 F 28.8
Max 272.7 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.0 F 0.0
5:00 224.9 F 0.0
10:00 213.9 F 0.0
15:00 205.7 F 0.0
20:00 201.7 F 0.0
25:00 203.0 F 0.0
30:00 209.1 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 24:23
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:02:35
CYCLE COMPLETE

01/03/2024 11:30 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____