

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3445

Name: RITTER 1  
Sterilizer ID: V2178198  
06/26/2023 12:17 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 191.2 F 0.0  
2:00 187.2 F 3.4  
4:00 206.7 F 7.9  
6:00 222.2 F 11.2  
8:00 230.9 F 9.2  
10:00 237.9 F 9.7  
12:00 243.3 F 12.7  
14:00 241.9 F 11.4  
16:00 255.1 F 18.1  
18:00 266.4 F 25.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.6 F 29.1  
2:00 271.5 F 28.3  
3:00 272.7 F 29.2  
4:00 272.0 F 28.6  
  
Min 271.4 F 28.3  
Max 272.7 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 209.4 F 0.0  
5:00 235.5 F 0.0  
10:00 235.2 F 0.0  
15:00 229.6 F 0.0  
20:00 228.1 F 0.0  
25:00 226.8 F 0.0  
30:00 226.4 F 0.0  
Drying Complete

FILLING: 2:00  
HEATING: 19:34  
STERILIZING: 4:00  
VENTING: 3:17  
DRYING: 30:00  
TOTAL CYCLE: 00:58:52  
CYCLE COMPLETE

06/26/2023 01:15 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_