
Midmark M11 - v1.0.0
Total Cycles: 1809

Name: RITTER 1
Sterilizer ID: V2178198
04/11/2022 11:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.6 F 0.0
2:00 182.9 F 4.2
4:00 191.5 F 9.1
6:00 213.9 F 7.4
8:00 228.1 F 8.5
10:00 239.9 F 12.2
12:00 239.7 F 10.7
14:00 236.3 F 8.8
16:00 250.8 F 15.6
18:00 262.7 F 22.3
20:00 272.5 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.1
1:00 272.5 F 28.9
2:00 271.7 F 28.3
3:00 271.4 F 28.5
4:00 272.5 F 28.9

Min 271.0 F 28.0
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.4 F 0.0
5:00 235.3 F 0.0
10:00 225.0 F 0.0
15:00 211.4 F 0.0
20:00 209.4 F 0.0
25:00 211.2 F 0.0
30:00 234.8 F 0.0
Drying Complete

FILLING: 1:23
HEATING: 20:16
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:48
CYCLE COMPLETE

04/11/2022 11:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____