
Midmark M11 - v1.0.0
Total Cycles: 4267

Name: RITTER 1
Sterilizer ID: V2178198
02/05/2024 09:17 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.6 F 0.0
2:00 77.0 F 0.2
4:00 87.9 F 1.0
6:00 114.2 F 3.0
8:00 146.6 F 5.7
10:00 169.8 F 8.8
12:00 191.2 F 12.3
14:00 210.4 F 3.2
16:00 220.4 F 6.3
18:00 228.2 F 9.7
20:00 237.0 F 13.4
22:00 233.9 F 8.8
24:00 244.4 F 13.5
26:00 241.3 F 10.9
28:00 252.6 F 16.5
30:00 262.4 F 22.4
32:00 272.0 F 28.6

STERILIZING
mm:ss Degrees PSI
0:00 272.3 F 28.4
1:00 272.8 F 28.9
2:00 272.2 F 29.0
3:00 271.5 F 28.4
4:00 272.5 F 28.7

Min 271.4 F 28.3
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 179.9 F 0.0
5:00 243.6 F 0.0
10:00 220.5 F 0.0
15:00 228.5 F 0.0
20:00 226.5 F 0.0
25:00 226.5 F 0.0
30:00 218.8 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 33:06
STERILIZING: 4:00
VENTING: 3:24
DRYING: 30:00
TOTAL CYCLE: 01:11:20
CYCLE COMPLETE

02/05/2024 10:28 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____