

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4468

Name: RITTER 1  
Sterilizer ID: V2178198  
04/17/2024 02:45 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.4 F 0.0  
2:00 143.3 F 1.9  
4:00 151.4 F 4.9  
6:00 180.7 F 8.6  
8:00 205.4 F 12.9  
10:00 216.2 F 5.3  
12:00 225.8 F 9.4  
14:00 238.0 F 11.1  
16:00 239.5 F 10.5  
18:00 235.2 F 8.2  
20:00 248.1 F 14.3  
22:00 258.6 F 20.4  
24:00 269.2 F 26.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.4  
1:00 272.0 F 28.9  
2:00 272.9 F 29.1  
3:00 271.5 F 28.5  
4:00 271.4 F 28.8

Min 271.0 F 28.3  
Max 273.2 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.6 F 0.0  
5:00 242.6 F 0.0  
10:00 222.6 F 0.0  
15:00 226.0 F 0.0  
20:00 224.8 F 0.0  
25:00 234.6 F 0.0  
30:00 230.3 F 0.0  
Drying Complete

FILLING: 1:02  
HEATING: 25:02  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:03:42  
CYCLE COMPLETE

04/17/2024 03:48 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_